



# SMSB

SENDI MAHIR SDN. BHD. (333138-T)

GST No: 001894494208

NO. 6, 8, 10 & 12 JALAN KAPAR 27/89, MEGAH INDUSTRIAL PARK,  
SEKSYEN 27, 40400 SHAH ALAM, SELANGOR DARUL EHSAN, MALAYSIA.

TEL: 03-5191 7388 (HUNTING LINE), 5191 7502, 5191 7592, 5192 9481 FAX: 03-5191 0675, 5191 9716  
EMAIL: enquiry@sendimahir.com ; marketing@sendimahir.com Website: www.sendimahir.com



MS ISO/IEC 17025  
CALIBRATION  
SAMM NO. 082

## CERTIFICATE OF CALIBRATION

Certificate No : SM18701100

Page 2 of 2 Pages

### Technical Information

Sensor Type : Infrared

Manufacturer Specification : N/A

Readability : 0.1/1 °C

Emissivity : 0.95

Calibration Distance : 400 mm

### Calibration Results :

All Unit In : °C

#### Accuracy Test

Nominal Value Source Emissivity ≈ 0.95	Correction	
	Before Adjustment	After Adjustment
50	+ 0.4	N/A
100	- 1	N/A
150	- 2	N/A
350	- 4	N/A

Measurement Uncertainty : ± 0.5 °C ( Range -20 to 100°C )  
 ± 0.8 °C ( Range 100 to 200°C )  
 ± 0.9 °C ( Range 200 to 400°C )

Note 1 : User Instrument Reading = Temperature Reading - Correction

Note 2 : To derive Temperature Reading = User Instrument Reading + Correction

Note 3 : Interpolation = Reading in between 2 test point may be derive by interpolate and plot a straight line graph where Temperature Reading(x-axis) Vs. Correction(y-axis).

Note 4 : Uncertainty = Parameter, associated with the result of measurement, that characterises the dispersion of the value that reasonably be attributed to the measurand.

Note 5 : If no adjustment was done refer to 'Correction before adjustment'. If adjustment was done refer to 'Correction after adjustment' to derive true value.

Note 6 : N/A = Not Available.